

ABL800 FLEX analyzer

Specifications



Measured parameters

Type	Parameter	Units	Measuring range
pH	pH**	pH scale	6.300–8.000
	cH ⁺	nmol/L	10.0–501
Blood gas	<i>pCO₂</i>	mmHg	5.0–250
		kPa	0.67–33.3
		Torr	5.0–250
Electrolyte	<i>pO₂</i>	mmHg	0.0–800
		kPa	0.00–107
		Torr	0.0–800
Metabolite	<i>cCl⁻</i>	mmol/L	7–350
		meq/L	7–350
	<i>cCa²⁺</i>	mmol/L	0.20–9.99
		meq/L	0.40–19.98
		mg/dL	0.80–40.04
	<i>cK⁺</i>	mmol/L	0.5–25.0
		meq/L	0.5–25.0
	<i>cNa⁺</i>	mmol/L	7–350
		meq/L	7–350
	<i>cGlu</i>	mmol/L	0.0–60
		mg/dL	0–1081
Oximetry	<i>cLac</i>	mmol/L	0.0–30
		mg/dL	0–270
		meq/L	0.0–30
	<i>cCrea</i>	μmol/L	10–1800
		mg/dL	0.11–20.4
	<i>ctBil</i>	μmol/L	0–1000
		mg/dL	0.0–58.5
		mg/L	0–585
	<i>ctHb</i>	g/dL	0.00–27.7
		mmol/L	0.00–17.2
		g/L	0.0–277
	<i>sO₂</i>	%	0.0–100.0
		Fraction	0.000–1.000
	<i>FO₂Hb</i>	%	0.0–100.0
		Fraction	0.000–1.000
	<i>FCOHb</i>	%	0.0–100.0
		Fraction	0.000–1.000
	<i>FMetHb</i>	%	0.0–100.0
		Fraction	0.000–1.000
	<i>FHHb</i>	%	0.0–100.0
		Fraction	0.000–1.000
	<i>FHbF</i>	%	0–100
		Fraction	0.00–1.00

** Also available as pH in pleural fluid.

The Measuring range for a parameter is the range within which the analyzer is physically capable of measuring.

The measuring range corresponds to the 'range of indication' as defined in the 'International vocabulary of basic and general terms in metrology' (VIM).

Derived parameters

pH(T)	cCa ²⁺ (pH=7.40)
cH ⁺ (T)	Anion Gap(K ⁺)
pCO ₂ (T)	Anion Gap
cHCO ₃ (P)	DO ₂
cBase(B)	Hct
cBase(B,ox)	pO ₂ (x)
cBase(Ecf)	pO ₂ (x,T)
cBase(Ecf,ox)	ctO ₂ (B)
cHCO ₃ (P,st)	ctO ₂ (a- \bar{v})
ctCO ₂ (P)	BO ₂
ctCO ₂ (B)	ctO ₂ (x)
pH(st)	FShunt
pO ₂ (T)	FShunt(T)
pO ₂ (A)	RI
pO ₂ (A,T)	RI(T)
p50	VO ₂
p50(T)	mOsm
p50(st)	Qx
pO ₂ (A-a)	Q _t
pO ₂ (A-a,T)	V(B)
pO ₂ (a/A)	sO ₂
pO ₂ (a/A,T)	FO ₂ Hb
pO ₂ (a)/FO ₂ (I)	FHHb
pO ₂ (a,T)/FO ₂ (I)	GFR, if AA
VCO ₂ /V(dry air)	GFR, if non AA

Parameters overview

Measuring Parameters							
pH & blood gases	Electrolytes	Metabolites			Oximetry		
pH, pCO ₂ , pO ₂	cNa ⁺ , cK ⁺ , cCa ²⁺ , cCl ⁻	cGlu,	cCrea	ctBil*	ctHb, sO ₂ * *	FMetHb, FCOHb, FO ₂ Hb, FHHb*	FHbF*
ABL805	✓	✓	✓				
ABL810	✓					✓	
ABL815	✓	✓	✓			✓	
ABL817	✓	✓	✓	✓		✓	
ABL820	✓					✓	✓
ABL825	✓	✓	✓			✓	✓
ABL827	✓	✓	✓	✓		✓	✓
ABL830	✓				✓	✓	✓
ABL835	✓	✓	✓		✓	✓	✓
ABL837	✓	✓	✓	✓	✓	✓	✓

* Measured by the spectrometer

Measuring system						
Analyzer		Mode	Sample volume	Measuring time (sec)	Cycle time (sec)	Throughput per hour
ABL825	ABL837					
✓		FLEXMODE (C)	35 – 195 µL	80 – 135	150 – 200	18 – 24
✓		all parameters (S)	195 µL	80	150	24
✓		all parameters, micro (S/C)	95 µL	135	200	18
	✓	all parameters (S)	250 µL	100	170	21
	✓	all parameters, micro (C)	125 µL	150	225	16
✓	✓	pH + BG + Oxi (S)	85 µL	80	170	21
✓	✓	pH + BG, Oxi, micro (C)	55 µL	100	170	21
✓	✓	Glu + Lac, micro (C)	35 µL	80	145	25
✓	✓	Oxi, micro (C)	35 µL	80	145	25
✓	✓	pH in pleural fluid (S)	85 µL	80	170	21
✓	✓	Expired air (S)	15 mL	80	170	21

Other analyzer versions will have other measuring times/volumes. S = Syringe C = Capillary

Hardware

Computer specifications

Intel® Athom™ Baytrail E3815
2 GB RAM
SSD
10.4" VGA captive color touch screen with USB

Interface

Integrated barcode reader
Serial line RS232
RJ45 Ethernet port
3 USB ports

FLEXQ

Module that allows queuing of samplers on the ABL800 FLEX.
Slots for samplers 3
Sampler type safePICO with safeTIPCAP
Sampler identification Integrated barcode scanner
Sample mixing time 8 seconds

Software

Software platform

Microsoft® Embedded System

Data capacity

Patient results: 2000
Calibration results: 1000
QC results: 1500
System messages and service registrations: 5000

Additional information

Dimensions

Width	71 cm	28 in
Height	57 cm	22 in
Depth	53 cm	21 in
Weight	37 kg	81 lbs

Data subject to change without notice.

Radiometer, the Radiometer logo, Connect & Care, ABL, AQT, TCM, AQURE, PICO, safePICO, safeCLINITUBES and QUALICHECK are trademarks of or used under license by Radiometer Medical ApS. © Radiometer Medical ApS, 2700 Brønshøj, Denmark, 2024. All Rights Reserved. MAPSS-000188 R4.

Calibration data

Automatic	Default interval	Interval options
1-point cal.	4 hours	after measurement, 30 min, 1, 2, 4 hours
2-point cal.	8 hours	after measurement, 1, 2, 4, 8 hours
1-point gas cal.*	2 hours	30 min, 1, 2 hours
System alignment	24 hours	
Cleaning	8 hours	8, 24 hours

Manual

tHb calibration 3 months never, 7 days, 1, 2, 3, 4, 6 months

* US only

Communication

Access to Local Area Network

Output protocols:

High-level protocols

POCT 1A

ASTM

HL7 (Version 2.2)

Low-level protocols

ASTM

Raw (serial only)

Transport layer

TCP/IP

RS232

Radiometer IT solution via Ethernet port

Other

Warm-up time Cold start: 29 min typical. Warm start: 5 min

Ambient temperature 15–32°C / 59–90°F

Relative humidity 20–80%

Thermostatting pH and blood gases,

37.0°C ± 0.15°C / 98.6°F ± 0.3°F

Electrolytes and metabolites,

37.0°C ± 0.25°C / 98.6°F ± 0.5°F

Spectrophotometer for measuring ctHb, sO₂, FHHb, FO₂Hb,

FCOHb, FMetHb, FHbF, ctBil on 128 wavelengths

Hemolyzer frequency 30 kHz intracuvette hemolysis

Barometer 450–800 mmHg

Power 100 – 240 V, 50/60 Hz, 270 VA